

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Report

for

San Lorenzo Valley Water District
13060 Highway 9
Boulder Creek, CA 95006-9119
Attention: Nate Gillespie
Fax: 831-338-8002

Date of Issue
03/08/2022

Pinda Seddos
EUROFINS EATON
ANALYTICAL, LLC



Utah ELCP CA00006

UMVN: Monica Van Natta
Project Manager

Report: 988302
Project: SLVWD
Group: VOC

* Accredited in accordance with TNI 2016 and ISO/IEC 17025:2017.

* Laboratory certifies that the test results meet all **TNI 2016 and ISO/IEC 17025:2017** requirements unless noted under the individual analysis.

* As applicable, this report consists of the cover page, State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms.

* Test results relate only to the sample(s) tested.

* Test results apply to the sample(s) as received, unless otherwise noted in the comments report (ISO/IEC 17025:2017).

* This report shall not be reproduced except in full, without the written approval of the laboratory.

* This report includes ISO/IEC 17025 and non-ISO 17025 accredited methods.

STATE CERTIFICATION LIST

State	Certification Number	State	Certification Number
Alabama	41060	Montana	Cert 0035
Arizona	AZ0778	Nebraska	NE-OS-21-13
Arkansas	CA00006	Nevada	CA00006
California	2813	New Hampshire *	2959
Colorado	CA00006	New Jersey *	CA 008
Connecticut	PH-0107	New Mexico	CA00006
Delaware	CA 006	New York *	11320
Florida *	E871024	North Carolina	06701
Georgia	947	North Dakota	R-009
Guam	21-008R	Ohio - 537.1	87786
Hawaii	CA00006	Oregon *	4034
Idaho	CA00006	Pennsylvania *	68-00565
Illinois	200033	Puerto Rico	CA00006
Indiana	C-CA-01	Rhode Island	LAO00326
Iowa – Asbestos	413	South Carolina	87016
Kansas *	E-10268	South Dakota	CA11320
Kentucky	90107	Tennessee	TN02839
Louisiana *	LA008	Texas *	T104704230-20-18
Maine	CA00006	Utah (Primary AB) *	CA00006
Maryland	224	Vermont	VT0114
Marianas Islands	MP0004	Virginia *	460260
Massachusetts	M-CA006	Washington	C838
Michigan	9906	EPA Region 5	CA00006
Mississippi	CA00006	Los Angeles County Sanitation Districts	10264

* NELAP/TNI Recognized Accreditation Bodies

ISO/IEC 17025:2917 Accredited Method List

The test listed below are accredited and met the requirements of ISO/IEC 17025 as verify by A2LA.

Refer to our certificates and scope of accreditations (no. 5890-1 and 5890-2) found at:

<https://www.eurofinsus.com/Eaton>

Test(s)	Method(s)	Potable Water *	Waste Water
Enterococci	Enterolert	x	x
<i>Escherichia coli</i> (Enumeration)	SM 9221 B.1 SM 9221 F	x	
Fecal Coliform (P/A and Enumeration)	SM 9221 C (MTF/EC), SM 9221 E (MTF/EC)	x	x
Fecal Streptococci and Enterococci	SM 9230 B	x	x
Heterotrophic Bacteria	SM 9215 B	x	
Legionella	Legiolert®	x	
<i>Pseudomonas aeruginosa</i>	Idexx Pseudalart	x	
Total Coliform (P/A and Enumeration)	SM 9221A, SM 9221B, SM 9221 C	x	x
Total Coliform, Total Coliform with Chlorine Present	SM 9221 B	x	x
Total Coliform/E. coli (P/A and Enumeration, Idexx Colilert, Idexx Colilert 18, Colisure)	SM 9223	x	
Total Microcystins and Nodularins	EPA 546	X	
Yeast and Mold	SM 9610	x	
1,2,3-Trichloropropane (TCP) at 5 PPT	CA SRL 524M-TCP	x	
1,4-Dioxane	EPA 522	x	
2,3,7,8-TCDD	Modified EPA 1613 B	x	
Acrylamide	* LCMS 2440)	x	
Algal Toxins/Microcystin	* LCMS 3570	x	
Alkalinity	SM 2320B	x	x
Ammonia	EPA 350.1, SM 4500-NH3 H		x
Asbestos	EPA 100.2	x	x
Bicarbonate Alkalinity as HCO3	SM 2330 B	x	x
BOD/CBOD	SM 5210 B		x
Bromate	* LCMS- 2447	x	
Carbonate as CO3	SM 2330 B	x	x
Carbonyls	EPA 556	x	x
Chemical Oxygen Demand	EPA 410.4, SM 5220D		x
Chlorinated Acids	EPA 515.4	x	
Chlorine Dioxide	Palin Test Chlordio X Plus, SM 4500-CLO2 D	x	
Chlorine, Free, Combined, Total Residual, Chloramines	SM 4500-CI G	x	
Color	SM2120B	x	
Conductivity	EPA 120.1, SM 2510B	x	x
Corrosivity (Langelier Index), Carbonate as CO3, Hydroxide as OH Calculated	SM 2330 B	x	
Cyanide (Amenable)	SM 4500-CN G	x	x
Cyanide (Free)	SM 4500CN F	x	x
Cyanide (Total)	EPA 335.4	x	x
Cyanogen Chloride (Screen)	* 335 Mod (WC-24467)	x	
Diquat and Paraquat	EPA 549.2	x	
DBP and HAA	SM 6251 B	x	
Dissolved Organic Carbon	SM 5310 C	x	
Dissolved Oxygen	SM 4500-O G		x
EDB/DCBP/TCP	EPA 504.1	x	
EDB/DBCP and Disinfection Byproducts	EPA 551.1	x	
EDTA and NTA	* WC-2454	x	
Endothall	EPA 548.1, *(LCMS-2445)	x	
Fluoride	SM 4500F C	x	x
Glyphosate	EPA 547	x	
Glyphosate and AMPA	* LCMS-3618	x	
Gross Alpha and Gross Beta	EPA 900.0	x	x

Test(s)	Method(s)	Potable Water *	Waste Water
Gross Alpha coprecipitation	SM 7110 C	x	x
Hardness	SM 2340 B	x	x
Hexavalent Chromium	EPA 218.6,	x	x
Hexavalent Chromium	EPA 218.7,	x	
Hexavalent Chromium	SM 3500-Cr B		x
Inorganic Anions and DBPs	EPA 300.0	x	x
Norganic Anions and DBPs	EPA 300.1	x	
Kjeldahl Nitrogen	EPA 351.2		x
Metals	EPA 200.7, EPA200.8	x	x
Nitrosamines	EEA-Agilent 521.1 (GCMS-24250)	x	
Nitrate/Nitrite Nitrogen	EPA 353.2	x	x
Odor	SM2150B	x	
Organohalide Pesticides and PCB	EPA 505	x	
Ortho Phosphate	SM 4500P E	x	
Oxyhalides Disinfection Byproducts	EPA 317.0	x	
Perchlorate	EPA 331.0	x	
Perchlorate (Low and High Levels)	EPA 314.0	x	
Perfluorinated Alkyl Acids	EPA 533, EPA 537, EPA 537.1	x	
PPCP and EDC	* LCMS-2443	x	
pH	EPA 150.1 SM 4500-H+ B	x	x
Phenolics – Low Level	*WC 2493 (EPA 420.2 and EPA 420.4 MOD)	x	x
Phenylurea Pesticides/Herbicides	* LCMS-2448	x	
Radium-226, Radium-228	GA Tech (Rad-2374)	x	
Radon-222	SM 7500RN	x	
Residue (Filterable)	SM 2540C	x	x
Residue (Non-Filterable)	SM 2540D		x
Residue (Total)	SM 2540B		x
Residue (Volatile)	EPA 160.4		x
Semi-Volatile Compounds	EPA 525.2	x	
Silica	SM 4500-SiO2 C	x	x
Sulfide	SM 4500-S D		x
Sulfite	SM 4500-SO3 B	x	x
Surfactants	SM 5540C	x	x
Taste and Odor	SM 6040 E	x	
Total Organic Carbon	SM 5310 C	x	x
Total Phenols	EPA 420.1		x
Total Phenols	EPA 420.4	x	x
Triazine Pesticides and their Degradates	* LCMS-3617	x	
Turbidity	EPA 180.1	x	x
Uranium by ICP/MS	EPA 200.8	x	
UV 254 Organic Constituents	SM 5910B	x	
VOCs	EPA 524.2	x	
VOCs	*(GCMS 2412) by EPA 524.2 modified	x	

(*) includes: Bottled Water, Drinking Water and Water as Component of Food & Beverage.

(+) In-House Method

Acknowledgement of Samples Received

Addr: **San Lorenzo Valley Water District**
 13060 Highway 9
 Boulder Creek, CA 95006-9119

Client ID: SANLORENZO

Folder #: 988302

Project: SLVWD

Sample Group: VOC

Attn: Nate Gillespie
 Phone: 831-216-9019

Project Manager: Monica Van Natta
 Phone: 559-326-4396

The following samples were received from you on **February 17, 2022 at 2158**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, LLC.

Sample #	Sample ID	Sample Date
<u>202202180208</u>	1000 Stewart Dr CLIP PS Code: CA4410014_DST_909 Sample Point Name: 1000 STEWART DR - VOCS - RIVERSIDE PZ Sample Type: RT @VOASDWAEDD_SB	02/16/2022 1120
<u>202202180209</u>	12788 Hwy 9 CLIP PS Code: CA4410014_DST_902 Sample Point Name: 12788 HWY 9 - VOCS - BIG STEEL ZONE Sample Type: RT Water Trax ID: 3BC98 @VOASDWAEDD_SB	02/16/2022 0930
<u>202202180210</u>	15030 Hwy 9 CLIP PS Code: CA4410014_DST_905 Sample Point Name: 15030 HWY 9 - VOCS - READER ZONE Sample Type: RT Water Trax ID: 3BCA5 @VOASDWAEDD_SB	02/16/2022 1010
<u>202202180211</u>	275 Brook Ln CLIP PS Code: CA4410014_DST_904 Sample Point Name: 275 BROOK LN - VOCS - LYON PRESSURE ZONE Sample Type: RT Water Trax ID: 3BCAB @VOASDWAEDD_SB	02/16/2022 0950
<u>202202180212</u>	801 Dry Well Rd CLIP PS Code: CA4410014_DST_908 Sample Point Name: 801 DRY WELL RD - VOCS - FIRE HYDRANT Sample Type: RT @VOASDWAEDD_SB	02/16/2022 1145
<u>202202180213</u>	814 Creek Drive CLIP PS Code: CA4410014_DST_907 Sample Point Name: 814 CREEK DR - VOCS - RIVERSIDE GROVE PZ Sample Type: RT Water Trax ID: 3BCA2 @VOASDWAEDD_SB	02/16/2022 1050
<u>202202180214</u>	815 Forest Drive CLIP PS Code: CA4410014_DST_901 Sample Point Name: 815 FOREST DR. - VOCS - SOUTH ZONE Sample Type: RT Water Trax ID: 3BCAE @VOASDWAEDD_SB	02/16/2022 0850
<u>202202180215</u>	Clear Creek and Alameda	02/16/2022 0910

Acknowledgement of Samples Received

Addr: **San Lorenzo Valley Water District**
 13060 Highway 9
 Boulder Creek, CA 95006-9119

Client ID: SANLORENZO

Folder #: 988302

Project: SLVWD

Sample Group: VOC

Attn: Nate Gillespie
 Phone: 831-216-9019

Project Manager: Monica Van Natta
 Phone: 559-326-4396

The following samples were received from you on **February 17, 2022** at **2158**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical, LLC.

Sample #	Sample ID	Sample Date
	CLIP PS Code: CA4410014_DST_903 Sample Point Name: CLEAR CREEK & ALAMEDA-VOCS-BIG STEEL PZ Sample Type: RT Water Trax ID: 3BC9B @VOASDWAEDD_SB	
<u>202202180216</u>	River and Crescent Dr CLIP PS Code: CA4410014_DST_906 Sample Point Name: RIVER & CRESCENT DR - VOCS -READER ZONE Sample Type: RT Water Trax ID: 3BCA3 @VOASDWAEDD_SB	02/16/2022 1030
<u>202202180217</u>	Riverside and Middle CLIP PS Code: CA4410014_DST_911 Sample Point Name: RIVERSIDE & MIDDLE - VOCS- BROOKDALE PZ Sample Type: RT Water Trax ID: 3BCB1 @VOASDWAEDD_SB	02/16/2022 0830
<u>202202180218</u>	Sylvan and Redwood CLIP PS Code: CA4410014_DST_910 Sample Point Name: SYLVAN & REDWOOD - VOCS - ECHO ZONE Sample Type: RT Water Trax ID: 3BC9F @VOASDWAEDD_SB	02/16/2022 1200
<u>202202180219</u>	Trip Blank C @VOASDWAEDD_SB TB	02/16/2022 0830

Test Description

@VOASDWAEDD_SB -- VOC EPA 524.2 Monrovia List

@VOASDWAEDD_SB TB -- VOC EPA 524.2 Monrovia List



Eaton Analytical

CHAIN OF CUSTODY RECORD

EUROFINS EATON ANALYTICAL USE ONLY:

750 Royal Oaks Drive, Suite 100
Monrovia, CA 91016-3629
Phone: 626 386 1100
Fax: 626 386 1101
800 566 LABS (800 566 5227)
Website: www.EatonAnalytical.com

LOGIN COMMENTS:

SAMPLE TEMP RECEIVED AT:

☐ Colton / No. California / Arizona
☒ Monrovia

CONDITION OF BLUE ICE:

METHOD OF SHIPMENT:

SAMPLES CHECKED AGAINST COC BY: 652

SAMPLES LOGGED IN BY:

SAMPLES REC'D DAY OF COLLECTION? ☐ (check for yes)

°C (Compliance: 4 ± 2 °C)

3.9 °C (Compliance: 4 ± 2 °C)

Frozen ☐ Partially Frozen ☒ Thawed ☐ No Ice ☐

Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: Top Line

TO BE COMPLETED BY SAMPLER:

COMPANY/AGENCY NAME: San Lorenzo Valley Water District		PROJECT CODE: 4410014	COMPLIANCE SAMPLES - Requires state forms ☒		NON-COMPLIANCE SAMPLES ☐	
EEA CLIENT CODE:		COC ID:	Type of samples (circle one): <u>ROUTINE</u> SPECIAL CONFIRMATION (eg. SDWA, Phase V, NPDES, FDA,...)		REGULATION INVOLVED:	

SAMPLE GROUP:		SEE ATTACHED BOTTLE ORDER FOR ANALYSES ☒ (check for yes), OR	
TAT requested: rush by adv notice only		list ANALYSES REQUIRED (enter number of bottles sent for each test for each sample)	

SAMPLE DATE	SAMPLE TIME	Client Lab ID	Sample ID (PS Codes)	MATRIX	Temp	Free Chlorine Residual	Total Chlorine	VOCs	SAMPLER COMMENTS WaterTrax #'s
2/16/22	1120	1000 Stewart Dr	CA4410014_DST_909	CFW	12.4	.90		X	N/A
2/16/22	0930	12788 Hwy 9	CA4410014_DST_902	CFW	10.6	1.52		X	3BC98
2/16/22	1010	15030 Hwy 9	CA4410014_DST_905	CFW	11.0	1.15		X	3BCA5
2/16/22	0950	275 Brook Lane	CA4410014_DST_904	CFW	10.1	1.36		X	3BCAB
2/16/22	1145	801 Dry Wells Rd	CA4410014_DST_908	CFW	12.6	.47		X	N/A
2/16/22	1050	814 Creek Dr	CA4410014_DST_907	CFW	10.3	1.10		X	3BCA2
2/16/22	0850	815 Forrest Dr	CA4410014_DST_901	CFW	10.1	.90		X	3BCAE
2/16/22	0910	Clear Creek & Alameda	CA4410014_DST_903	CFW	10.4	1.42		X	3BC9B
2/16/22	1030	River & Crescent Dr	CA4410014_DST_906	CFW	10.0	1.39		X	3BCA3
2/16/22	0830	Riverside and Middle Dr	CA4410014_DST_911	CFW	11.4	1.40		X	3BCB1

* MATRIX TYPES: RSW = Raw Surface Water CFW = Chlor(am)inated Finished Water SEAW = Sea Water BW = Bottled Water SO = Soil
RGW = Raw Ground Water FW = Other Finished Water WW = Waste Water SW = Storm Water SL = Sludge O = Other - Please Identify

SAMPLED BY:	SIGNATURE	PRINT NAME	COMPANY/TITLE	DATE	TIME
RELINQUISHED BY:	<u>Kraig Mitchell</u>	Kraig Mitchell	San Lorenzo Valley Water District	2/16/22	0830
RECEIVED BY:	<u>Kraig Mitchell</u>	Kraig Mitchell	San Lorenzo Valley Water District	2/16/22	1530
RELINQUISHED BY:	<u>Gustavo Sanchez</u>	Gustavo Sanchez	EEA	2/17/22	2158
RECEIVED BY:					

Factor Analytical

INTERNAL CHAIN OF CUSTODY RECORD

EEA Folder Number:

Note: If examples are out of temperature range, let the ASMs know. ASMs will determine whether to proceed with analysis or not.

SAMPLES REC'D DAY OF COLLECTION?		Yes / No
1	2	
3	4	
5	6	
7	8	
9	10	
11	12	
13	14	
15	16	
17	18	
19	20	
21	22	
23	24	
25	26	
27	28	
29	30	
31	32	
33	34	
35	36	
37	38	
39	40	
41	42	
43	44	
45	46	
47	48	
49	50	
51	52	
53	54	
55	56	
57	58	
59	60	
61	62	
63	64	
65	66	
67	68	
69	70	
71	72	
73	74	
75	76	
77	78	
79	80	
81	82	
83	84	
85	86	
87	88	
89	90	
91	92	
93	94	
95	96	
97	98	
99	100	

IP Gun ID = 413A (Observation = 4.0 °C) (Corr. Factor = -0.1 °C) (Final = 3.9 °C)

Source of Ice: Pool	Synthetic	No Ice	CONDITION OF ICE:	Frozen	Partially Frozen	Thawed	N/A
	☒	☐		☐	☒	☐	☐

METHOD OF SHIPMENT: Pick-Up / Walk-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: _____

Compliance Acceptance Criteria:

- 1) Chemistry: >0, $\leq 8^{\circ}\text{C}$, not frozen (NELAP) (if received after 24 hrs of sample collection)
- 2) Microbiology, Distribution: < 10°C , not frozen (can be $\geq 10^{\circ}\text{C}$ if received on ice the same day as sample collection, within 8 hours)
- 3) Microbiology, Surface Water: < 10°C (if received after 2 hours of sample collection)

If out of temperature range for both Chemistry and Microbiology samples and temperature does not confirm, then measure the temperature of each quadrant and record each temperature of the quadrants.

1 = Observation = _____ °C Corr. Factor = _____ °C Final = _____ °C	2 = Observation = _____ °C Corr. Factor = _____ °C Final = _____ °C
3 = Observation = _____ °C Corr. Factor = _____ °C Final = _____ °C	4 = Observation = _____ °C Corr. Factor = _____ °C Final = _____ °C

- 4 Dioxin (1613 or 2,3,7,8 TCDD): must be between 0-4 °C, not frozen (If received after 24 hrs of sample collection)
- 5) pH Check. Manufacturer: _____ Lot Number: _____ pH strip type: 0 - 14 or _____ Expiration Date: _____ Results: _____
- 6) Chlorine check. Manufacturer: Sansafe. Lot No.: _____ Expiration Date: _____ Results: _____

VOA and Radon

No Samples with Headspace:

Samples with Headspace (see below):

Use additional COFC for additional bottles)

Headspace Documentation (use additional Vials and Raoult Internal Vials for additional bottles). International clients: CMS methods use 40 mL vials.

Exempt from headspace concerns: Methods 515.4, HAA(9241,552), 505, SPME, @CH, 532LCMS, 556, 536, Anatoxin, LCMS methods using 40 ml vials, international standards.	None/<6	Samp ID	Bottle #	Test	Test
	None/<6			>6mm	>6mm

[illegible]

Note Sample IDs which have dissimilar headspace (i.e. potential sampling errors):

SIGNATURE

PRINT NAME

COMPANY/TITLE

DATE _____

TIME

RECEIVED BY:

OBJ:	<i>[Signature]</i>	<i>[Signature]</i>
	<i>[Signature]</i>	<i>[Signature]</i>

Eurofins Eaton Analytical


22-71-20 21588

SIGNATURE

PRINT NAME

COMPANY/TITLE

TIME



Eaton Analytical

Kit Order for San Lorenzo Valley Water District

Monica Van Natta is your Eurofins Eaton Analytical, LLC Service Manager

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
(626) 386-1100 FAX (866) 988-3757

Kit #: 309347



Created By: - [AutoGenerated]
Deliver By: 02/01/2022
STG: Bottle Orders
Ice Type: G

Client ID: SANLORENZO



Project Code: SLWWD Bottle Orders

Group Name: VOC

PO#/JOB#:

Description: Every 3 months on 1

Note: Sampler Please return this paper with your samples

Created Date & Time: 1/2/2022 12:07:25AM

Ship Sample Kits to
San Lorenzo Valley Water District
195 Kirby Street
Felton, CA 95018

Attn: Shipping - Water Testing Bottles
Phone: 831-216-9019

Send Report to
San Lorenzo Valley Water District
13060 Highway 9
Boulder Creek, CA 95006-9119

Attn: Nate Gillespie
Phone: 831-216-9019
Fax: 831-338-8002

Billing Address
San Lorenzo Valley Water District
13060 Highway 9
Boulder Creek, 95006

Attn: Rachel Munoz
Phone: 831.338.2153
Fax: 831.338.7986

# of Sample Tests	Bottle Qty - Type [preservative information]	Total	UN DOT #
1 @VOASDWA TB C	2 - 40ml amber glass vial [2hmg AA+ H2O+10 drop 1:1 HCL]	2	UN1789
11 @VOASDWA C	3 - 40ml amber glass vial [2hmg Ascorbic+drop 2ml 1:1 HCL]	33	UN1789

Sum Tests: 12

Sum Bottles: 35

Comments

Code

Status

Date Shipped

Via

Tracking #

of Coolers

Prenared Bv

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Comments

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Folder Comments

Analytical results for Volatiles by 524.2 are submitted by Eurofins Eaton Analytical in
South Bend IN CA cert 2920 exp 6-30-22

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Hits

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Analyzed	Analyte	Sample ID	Result	Federal MCL	Units	MRL
202202180208 <u>1000 Stewart Dr</u>						
02/25/2022 16:03	Bromodichloromethane		10		ug/L	0.5
02/25/2022 16:03	Chloroform		21		ug/L	0.5
02/25/2022 16:03	Dibromochloromethane		4.9		ug/L	0.5
02/25/2022 16:03	Total THMs		35.9		ug/L	0.5
202202180209 <u>12788 Hwy 9</u>						
02/25/2022 16:26	Bromodichloromethane		4.9		ug/L	0.5
02/25/2022 16:26	Chloroform		6.3		ug/L	0.5
02/25/2022 16:26	Dibromochloromethane		3		ug/L	0.5
02/25/2022 16:26	Total THMs		14.2		ug/L	0.5
202202180210 <u>15030 Hwy 9</u>						
02/25/2022 16:50	Bromodichloromethane		5.4		ug/L	0.5
02/25/2022 16:50	Chloroform		6.9		ug/L	0.5
02/25/2022 16:50	Dibromochloromethane		3.2		ug/L	0.5
02/25/2022 16:50	Total THMs		15.5		ug/L	0.5
202202180211 <u>275 Brook Ln</u>						
02/25/2022 17:13	Bromodichloromethane		8.1		ug/L	0.5
02/25/2022 17:13	Chloroform		12		ug/L	0.5
02/25/2022 17:13	Dibromochloromethane		4.6		ug/L	0.5
02/25/2022 17:13	Total THMs		24.7		ug/L	0.5
202202180212 <u>801 Dry Well Rd</u>						
02/25/2022 17:36	Bromodichloromethane		12		ug/L	0.5
02/25/2022 17:36	Bromoform		0.51		ug/L	0.5
02/25/2022 17:36	Chloroform		28		ug/L	0.5
02/25/2022 17:36	Dibromochloromethane		5.1		ug/L	0.5
02/25/2022 17:36	Total THMs		45.6		ug/L	0.5
202202180213 <u>814 Creek Drive</u>						
02/25/2022 17:59	Bromodichloromethane		9.6		ug/L	0.5
02/25/2022 17:59	Chloroform		19		ug/L	0.5
02/25/2022 17:59	Dibromochloromethane		4.2		ug/L	0.5
02/25/2022 17:59	Total THMs		32.8		ug/L	0.5
202202180214 <u>815 Forest Drive</u>						
02/25/2022 18:23	Bromodichloromethane		5.9		ug/L	0.5
02/25/2022 18:23	Chloroform		7.5		ug/L	0.5

SUMMARY OF POSITIVE DATA ONLY

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Analyzed	Analyte	Sample ID	Result	Federal MCL	Units	MRL
02/25/2022 18:23	Dibromochloromethane		3.4		ug/L	0.5
02/25/2022 18:23	Total THMs		16.8		ug/L	0.5
	202202180215	<u>Clear Creek and Alameda</u>				
02/25/2022 18:46	Bromodichloromethane		5.8		ug/L	0.5
02/25/2022 18:46	Chloroform		7.4		ug/L	0.5
02/25/2022 18:46	Dibromochloromethane		3.4		ug/L	0.5
02/25/2022 18:46	Total THMs		16.6		ug/L	0.5
	202202180216	<u>River and Crescent Dr</u>				
02/25/2022 19:09	Bromodichloromethane		6.9		ug/L	0.5
02/25/2022 19:09	Chloroform		9.5		ug/L	0.5
02/25/2022 19:09	Dibromochloromethane		3.8		ug/L	0.5
02/25/2022 19:09	Total THMs		20.2		ug/L	0.5
	202202180217	<u>Riverside and Middle</u>				
02/25/2022 19:32	Bromodichloromethane		7.3		ug/L	0.5
02/25/2022 19:32	Chloroform		9.4		ug/L	0.5
02/25/2022 19:32	Dibromochloromethane		4.2		ug/L	0.5
02/25/2022 19:32	Total THMs		20.9		ug/L	0.5
	202202180218	<u>Sylvan and Redwood</u>				
02/25/2022 19:55	Bromodichloromethane		9.8		ug/L	0.5
02/25/2022 19:55	Chloroform		22		ug/L	0.5
02/25/2022 19:55	Dibromochloromethane		4.7		ug/L	0.5
02/25/2022 19:55	Total THMs		36.5		ug/L	0.5

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
1000 Stewart Dr (202202180208)						Sampled on 02/16/2022 1120			
Sample Type: RT									
EPA 524.2 - VOC EPA 524.2 Monrovia List									
02/25/22	02/25/22 16:03			(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Bromodichloromethane	10	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Chloroform	21	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 16:03			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Dibromochloromethane	4.9	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 16:03			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Total THMs	35.9	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:03			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 16:03			(EPA 524.2)	1,2-Dichlorobenzene-d4	78	%		1

12788 Hwy 9 (202202180209)

Sampled on 02/16/2022 0930

Rounding on totals after summation.
(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
Sample Type: RT									
Water Trax ID: 3BC98									
EPA 524.2 - VOC EPA 524.2 Monrovia List									
02/25/22	02/25/22 16:26			(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Bromodichloromethane	4.9	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Chloroform	6.3	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1

Rounding on totals after summation.
(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 16:26			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Dibromochloromethane	3	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 16:26			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Total THMs	14.2	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:26			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 16:26			(EPA 524.2)	1,2-Dichlorobenzene-d4	77	%		1

15030 Hwy 9 (202202180210)

Sampled on 02/16/2022 1010

Sample Type: RT
Water Trax ID: 3BCA5

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (626) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
EPA 524.2 - VOC EPA 524.2 Monrovia List									
02/25/22	02/25/22 16:50			(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Bromodichloromethane	5.4	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Chloroform	6.9	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 16:50			(EPA 524.2)	Dibromochloromethane	3.2	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 16:50			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Total THMs	15.5	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 16:50			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 16:50			(EPA 524.2)	1,2-Dichlorobenzene-d4	79	%		1

275 Brook Ln (202202180211)

Sampled on 02/16/2022 0950

Sample Type: RT
Water Trax ID: 3BCAB

EPA 524.2 - VOC EPA 524.2 Monrovia List

Rounding on totals after summation.
(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (626) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 17:13			(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Bromodichloromethane	8.1	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Chloroform	12	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Dibromochloromethane	4.6	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 17:13			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 17:13			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Total THMs	24.7	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:13			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 17:13			(EPA 524.2)	1,2-Dichlorobenzene-d4	75	%		1

801 Dry Well Rd (202202180212)

Sampled on 02/16/2022 1145

Sample Type: RT

EPA 524.2 - VOC EPA 524.2 Monrovia List

02/25/22	02/25/22 17:36	(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36	(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (626) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 17:36			(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Bromodichloromethane	12	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Bromoform	0.51	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Chloroform	28	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Dibromochloromethane	5.1	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 17:36			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 17:36			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Total THMs	45.6	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:36			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 17:36			(EPA 524.2)	1,2-Dichlorobenzene-d4	78	%		1

814 Creek Drive (202202180213)

Sampled on 02/16/2022 1050

Sample Type: RT
Water Trax ID: 3BCA2

EPA 524.2 - VOC EPA 524.2 Monrovia List

02/25/22	02/25/22 17:59	(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59	(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59	(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 17:59			(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Bromodichloromethane	9.6	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Chloroform	19	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Dibromochloromethane	4.2	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 17:59			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 17:59			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Total THMs	32.8	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 17:59			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 17:59			(EPA 524.2)	1,2-Dichlorobenzene-d4	77	%		1

815 Forest Drive (202202180214)

Sampled on 02/16/2022 0850

Sample Type: RT
Water Trax ID: 3BCAE

EPA 524.2 - VOC EPA 524.2 Monrovia List

02/25/22	02/25/22 18:23	(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23	(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23	(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23	(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (626) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 18:23			(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Bromodichloromethane	5.9	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Chloroform	7.5	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Dibromochloromethane	3.4	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 18:23			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 18:23			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Total THMs	16.8	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 18:23			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 18:23			(EPA 524.2)	1,2-Dichlorobenzene-d4	75	%		1

Clear Creek and Alameda (202202180215)

Sampled on 02/16/2022 0910

Sample Type: RT
Water Trax ID: 3BC9B

EPA 524.2 - VOC EPA 524.2 Monrovia List

02/25/22	02/25/22 18:46	(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46	(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46	(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46	(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46	(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 18:46			(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Bromodichloromethane	5.8	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Chloroform	7.4	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Dibromochloromethane	3.4	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 18:46			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 18:46			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Total THMs	16.6	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 18:46			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 18:46			(EPA 524.2)	1,2-Dichlorobenzene-d4	77	%		1

River and Crescent Dr (202202180216)

Sampled on 02/16/2022 1030

Sample Type: RT
Water Trax ID: 3BCA3

EPA 524.2 - VOC EPA 524.2 Monrovia List

02/25/22	02/25/22 19:09	(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09	(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09	(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09	(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09	(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09	(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (626) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 19:09			(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Bromodichloromethane	6.9	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Chloroform	9.5	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Dibromochloromethane	3.8	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (626) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 19:09			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 19:09			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Total THMs	20.2	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:09			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 19:09			(EPA 524.2)	1,2-Dichlorobenzene-d4	77	%		1

Riverside and Middle (202202180217)

Sampled on 02/16/2022 0830

Sample Type: RT
Water Trax ID: 3BCB1

EPA 524.2 - VOC EPA 524.2 Monrovia List

02/25/22	02/25/22 19:32	(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32	(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32	(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32	(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32	(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32	(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32	(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 19:32			(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Bromodichloromethane	7.3	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Chloroform	9.4	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Dibromochloromethane	4.2	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 19:32			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 19:32			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Total THMs	20.9	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:32			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 19:32			(EPA 524.2)	1,2-Dichlorobenzene-d4	77	%		1

Sylvan and Redwood (202202180218)

Sampled on 02/16/2022 1200

Sample Type: RT
Water Trax ID: 3BC9F

EPA 524.2 - VOC EPA 524.2 Monrovia List

02/25/22	02/25/22 19:55	(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55	(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55	(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55	(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55	(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55	(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55	(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55	(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (626) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 19:55			(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Bromodichloromethane	9.8	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Chloroform	22	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Dibromochloromethane	4.7	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 19:55			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 19:55			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Total THMs	36.5	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 19:55			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 19:55			(EPA 524.2)	1,2-Dichlorobenzene-d4	76	%		1

Trip Blank C (202202180219)

Sampled on 02/16/2022 0830

EPA 524.2 - VOC EPA 524.2 Monrovia List

02/25/22	02/25/22 20:18	(EPA 524.2)	1,1,1,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18	(EPA 524.2)	1,1,1-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18	(EPA 524.2)	1,1,2,2-Tetrachloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18	(EPA 524.2)	1,1,2-Trichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18	(EPA 524.2)	1,1-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18	(EPA 524.2)	1,1-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18	(EPA 524.2)	1,1-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18	(EPA 524.2)	1,2,3-Trichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18	(EPA 524.2)	1,2,3-Trichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18	(EPA 524.2)	1,2,4-Trichlorobenzene	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 20:18			(EPA 524.2)	1,2,4-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	1,2-Dichloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	1,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	1,3,5-Trimethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	1,3-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	2,2-Dichloropropane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	2-Butanone (MEK)	ND	ug/L	5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	2-Chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	4-chlorotoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	4-Isopropyltoluene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	4-methyl-2-Pentanone	ND	ug/L	2	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Benzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Bromobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Bromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Bromodichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Bromoethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Bromoform	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Bromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Carbon disulfide	ND	ug/L	5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Carbon Tetrachloride	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Chlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Chloroethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Chloroform	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Chloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	cis-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	cis-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Dibromochloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Dibromomethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Dichlorodifluoromethane (Freon 12)	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Dichloromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Diisopropyl Ether (DIPE)	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Ethylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Ethyl-t-Butyl Ether (ETBE)	ND	ug/L	3	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Hexachlorobutadiene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Isopropylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	m,p-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	m-Dichlorobenzene	ND	ug/L	0.5	1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

Tel: (626) 386-1100
Fax: (866) 988-3757
1 800 566 LABS (1 800 566 5227)

Laboratory Data

Report: 988302
Project: SLVWD
Group: VOC

San Lorenzo Valley Water District
Nate Gillespie
13060 Highway 9
Boulder Creek, CA 95006-9119

Samples Received on:
02/17/2022 2158

Prepped	Analyzed	Prep Batch	Analytical Batch	Method	Analyte	Result	Units	MRL	Dilution
02/25/22	02/25/22 20:18			(EPA 524.2)	Methyl Tert-butyl ether (MTBE)	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Naphthalene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	n-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	n-Propylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	o-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	o-Xylene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	p-Dichlorobenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	sec-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Styrene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Tert-Amyl-Methyl Ether (TAME)	ND	ug/L	3	1
02/25/22	02/25/22 20:18			(EPA 524.2)	tert-Butylbenzene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Tetrachloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Toluene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Total 1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Total THMs	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Total Xylenes	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	trans-1,2-Dichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	trans-1,3-Dichloropropene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Trichloroethylene	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Trichlorofluoromethane	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Trichlorotrifluoroethane (Freon 113)	ND	ug/L	0.5	1
02/25/22	02/25/22 20:18			(EPA 524.2)	Vinyl chloride	ND	ug/L	0.2	1
02/25/22	02/25/22 20:18			(EPA 524.2)	1,2-Dichlorobenzene-d4	79	%		1

Rounding on totals after summation.

(c) - indicates calculated results. Analysis is a calculated result. Reported results are not rounded until the final step before reporting. Therefore methods that use a test result with further calculation may have slight differences in final result than the component analyses.

ANALYTICAL REPORT

Eurofins Eaton South Bend
110 S Hill Street
South Bend, IN 46617
Tel: (574)233-4777

Laboratory Job ID: 810-15928-1
Client Project/Site: 988302

For:
Eurofins Eaton Analytical
750 Royal Oaks Drive
Suite 100
Monrovia, California 91016

Attn: Jaclyn Contreras



Authorized for release by:
3/7/2022 3:47:50 PM
Neely Davis, Project Manager
(574)233-4777

neely.davis@eurofinset.com

Designee for
Karen Fullmer, Project Manager
(574)233-4777
karen.fullmer@eurofinset.com

LINKS

Review your project
results through

TotalAccess

Have a Question?



Visit us at:

www.eurofinsus.com/Env

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.

Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Detection Summary	5
Client Sample Results	7
Surrogate Summary	27
QC Sample Results	28
QC Association Summary	30
Lab Chronicle	31
Certification Summary	33
Method Summary	34
Sample Summary	35
Chain of Custody	36
Receipt Checklists	39



Definitions/Glossary

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
□	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Job ID: 810-15928-1

Laboratory: Eurofins Eaton South Bend

Narrative

Job Narrative
810-15928-1

Comments

No additional comments.

Receipt

The samples were received on 2/24/2022 10:15 AM. Unless otherwise noted below, the samples arrived in good condition, and where required, properly preserved and on ice. The temperature of the cooler at receipt was 2.0° C.

GC/MS VOA

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

Detection Summary

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180208

Lab Sample ID: 810-15928-1

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	35.90		0.5000	ug/L	1		524.2	Total/NA
Bromodichloromethane	10		0.50	ug/L	1		524.2	Total/NA
Chloroform	21		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	4.9		0.50	ug/L	1		524.2	Total/NA

Client Sample ID: 202202180209

Lab Sample ID: 810-15928-2

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	14.20		0.5000	ug/L	1		524.2	Total/NA
Bromodichloromethane	4.9		0.50	ug/L	1		524.2	Total/NA
Chloroform	6.3		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	3.0		0.50	ug/L	1		524.2	Total/NA

Client Sample ID: 202202180210

Lab Sample ID: 810-15928-3

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	15.50		0.5000	ug/L	1		524.2	Total/NA
Bromodichloromethane	5.4		0.50	ug/L	1		524.2	Total/NA
Chloroform	6.9		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	3.2		0.50	ug/L	1		524.2	Total/NA

Client Sample ID: 202202180211

Lab Sample ID: 810-15928-4

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	24.70		0.5000	ug/L	1		524.2	Total/NA
Bromodichloromethane	8.1		0.50	ug/L	1		524.2	Total/NA
Chloroform	12		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	4.6		0.50	ug/L	1		524.2	Total/NA

Client Sample ID: 202202180212

Lab Sample ID: 810-15928-5

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	45.6		0.5000	ug/L	1		524.2	Total/NA
Bromoform	0.51		0.50	ug/L	1		524.2	Total/NA
Bromodichloromethane	12		0.50	ug/L	1		524.2	Total/NA
Chloroform	28		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	5.1		0.50	ug/L	1		524.2	Total/NA

Client Sample ID: 202202180213

Lab Sample ID: 810-15928-6

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	32.80		0.5000	ug/L	1		524.2	Total/NA
Bromodichloromethane	9.6		0.50	ug/L	1		524.2	Total/NA
Chloroform	19		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	4.2		0.50	ug/L	1		524.2	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Eaton South Bend

Detection Summary

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180214

Lab Sample ID: 810-15928-7

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	16.80		0.5000	ug/L	1		524.2	Total/NA
Bromodichloromethane	5.9		0.50	ug/L	1		524.2	Total/NA
Chloroform	7.5		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	3.4		0.50	ug/L	1		524.2	Total/NA

Client Sample ID: 202202180215

Lab Sample ID: 810-15928-8

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	16.60		0.5000	ug/L	1		524.2	Total/NA
Bromodichloromethane	5.8		0.50	ug/L	1		524.2	Total/NA
Chloroform	7.4		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	3.4		0.50	ug/L	1		524.2	Total/NA

Client Sample ID: 202202180216

Lab Sample ID: 810-15928-9

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	20.20		0.5000	ug/L	1		524.2	Total/NA
Bromodichloromethane	6.9		0.50	ug/L	1		524.2	Total/NA
Chloroform	9.5		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	3.8		0.50	ug/L	1		524.2	Total/NA

Client Sample ID: 202202180217

Lab Sample ID: 810-15928-10

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	20.90		0.5000	ug/L	1		524.2	Total/NA
Bromodichloromethane	7.3		0.50	ug/L	1		524.2	Total/NA
Chloroform	9.4		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	4.2		0.50	ug/L	1		524.2	Total/NA

Client Sample ID: 202202180218

Lab Sample ID: 810-15928-11

PWSID Number: CA4410014

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Trihalomethanes, Total	36.50		0.5000	ug/L	1		524.2	Total/NA
Bromodichloromethane	9.8		0.50	ug/L	1		524.2	Total/NA
Chloroform	22		0.50	ug/L	1		524.2	Total/NA
Dibromochloromethane	4.7		0.50	ug/L	1		524.2	Total/NA

Client Sample ID: 202202180219

Lab Sample ID: 810-15928-12

PWSID Number: CA4410014

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180208

Lab Sample ID: 810-15928-1

Date Collected: 02/16/22 11:20

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	35.90		0.5000	ug/L			02/25/22 16:03	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 16:03	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 16:03	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 16:03	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 16:03	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 16:03	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 16:03	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 16:03	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 16:03	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 16:03	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 16:03	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 16:03	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 16:03	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 16:03	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 16:03	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 16:03	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 16:03	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 16:03	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 16:03	1
Benzene	<0.50		0.50	ug/L			02/25/22 16:03	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
Bromoform	<0.50		0.50	ug/L			02/25/22 16:03	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 16:03	1
Bromodichloromethane	10		0.50	ug/L			02/25/22 16:03	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 16:03	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 16:03	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 16:03	1
Chloroform	21		0.50	ug/L			02/25/22 16:03	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 16:03	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 16:03	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 16:03	1
Dibromochloromethane	4.9		0.50	ug/L			02/25/22 16:03	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 16:03	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 16:03	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 16:03	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 16:03	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 16:03	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180208

Lab Sample ID: 810-15928-1

Date Collected: 02/16/22 11:20

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 16:03	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 16:03	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 16:03	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
Styrene	<0.50		0.50	ug/L			02/25/22 16:03	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 16:03	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 16:03	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 16:03	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 16:03	1
Toluene	<0.50		0.50	ug/L			02/25/22 16:03	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 16:03	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 16:03	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 16:03	1
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 16:03	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 16:03	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 16:03	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 16:03	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 16:03	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 16:03	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	78		70 - 130		02/25/22 16:03	1
1,2-Dichloroethane-d4 (Surr)	103		70 - 130		02/25/22 16:03	1
4-Bromofluorobenzene (Surr)	84		70 - 130		02/25/22 16:03	1
Toluene-d8 (Surr)	100		70 - 130		02/25/22 16:03	1

Client Sample ID: 202202180209

Lab Sample ID: 810-15928-2

Date Collected: 02/16/22 09:30

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	14.20		0.5000	ug/L			02/25/22 16:26	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 16:26	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 16:26	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 16:26	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 16:26	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 16:26	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 16:26	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 16:26	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 16:26	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 16:26	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 16:26	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180209

Lab Sample ID: 810-15928-2

Date Collected: 02/16/22 09:30

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 16:26	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 16:26	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 16:26	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 16:26	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 16:26	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 16:26	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 16:26	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 16:26	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 16:26	1
Benzene	<0.50		0.50	ug/L			02/25/22 16:26	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
Bromoform	<0.50		0.50	ug/L			02/25/22 16:26	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 16:26	1
Bromodichloromethane	4.9		0.50	ug/L			02/25/22 16:26	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 16:26	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 16:26	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 16:26	1
Chloroform	6.3		0.50	ug/L			02/25/22 16:26	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 16:26	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 16:26	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 16:26	1
Dibromochloromethane	3.0		0.50	ug/L			02/25/22 16:26	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 16:26	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 16:26	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 16:26	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 16:26	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 16:26	1
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 16:26	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 16:26	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 16:26	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
Styrene	<0.50		0.50	ug/L			02/25/22 16:26	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 16:26	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 16:26	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 16:26	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 16:26	1
Toluene	<0.50		0.50	ug/L			02/25/22 16:26	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 16:26	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 16:26	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 16:26	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180209

Lab Sample ID: 810-15928-2

Date Collected: 02/16/22 09:30

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 16:26	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 16:26	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 16:26	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 16:26	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 16:26	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 16:26	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	77		70 - 130		02/25/22 16:26	1
1,2-Dichloroethane-d4 (Surr)	104		70 - 130		02/25/22 16:26	1
4-Bromofluorobenzene (Surr)	80		70 - 130		02/25/22 16:26	1
Toluene-d8 (Surr)	100		70 - 130		02/25/22 16:26	1

Client Sample ID: 202202180210

Lab Sample ID: 810-15928-3

Date Collected: 02/16/22 10:10

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	15.50		0.5000	ug/L			02/25/22 16:50	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 16:50	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 16:50	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 16:50	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 16:50	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 16:50	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 16:50	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 16:50	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 16:50	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 16:50	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 16:50	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 16:50	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 16:50	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 16:50	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 16:50	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 16:50	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 16:50	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 16:50	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 16:50	1
Benzene	<0.50		0.50	ug/L			02/25/22 16:50	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 16:50	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180210

Lab Sample ID: 810-15928-3

Date Collected: 02/16/22 10:10

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Bromoform	<0.50		0.50	ug/L			02/25/22 16:50	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 16:50	1
Bromodichloromethane	5.4		0.50	ug/L			02/25/22 16:50	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 16:50	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 16:50	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 16:50	1
Chloroform	6.9		0.50	ug/L			02/25/22 16:50	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 16:50	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 16:50	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 16:50	1
Dibromochloromethane	3.2		0.50	ug/L			02/25/22 16:50	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 16:50	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 16:50	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 16:50	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 16:50	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 16:50	1
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 16:50	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 16:50	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 16:50	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
Styrene	<0.50		0.50	ug/L			02/25/22 16:50	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 16:50	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 16:50	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 16:50	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 16:50	1
Toluene	<0.50		0.50	ug/L			02/25/22 16:50	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 16:50	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 16:50	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 16:50	1
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 16:50	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 16:50	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 16:50	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 16:50	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 16:50	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 16:50	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	79		70 - 130		02/25/22 16:50	1
1,2-Dichloroethane-d4 (Surr)	106		70 - 130		02/25/22 16:50	1
4-Bromofluorobenzene (Surr)	81		70 - 130		02/25/22 16:50	1
Toluene-d8 (Surr)	100		70 - 130		02/25/22 16:50	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180211

Lab Sample ID: 810-15928-4

Date Collected: 02/16/22 09:50

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	24.70		0.5000	ug/L			02/25/22 17:13	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 17:13	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 17:13	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 17:13	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 17:13	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 17:13	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 17:13	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 17:13	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 17:13	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 17:13	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 17:13	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 17:13	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 17:13	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 17:13	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 17:13	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 17:13	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 17:13	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 17:13	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 17:13	1
Benzene	<0.50		0.50	ug/L			02/25/22 17:13	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
Bromoform	<0.50		0.50	ug/L			02/25/22 17:13	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 17:13	1
Bromodichloromethane	8.1		0.50	ug/L			02/25/22 17:13	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 17:13	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 17:13	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 17:13	1
Chloroform	12		0.50	ug/L			02/25/22 17:13	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 17:13	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 17:13	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 17:13	1
Dibromochloromethane	4.6		0.50	ug/L			02/25/22 17:13	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 17:13	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 17:13	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 17:13	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 17:13	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 17:13	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180211

Lab Sample ID: 810-15928-4

Date Collected: 02/16/22 09:50

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 17:13	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 17:13	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 17:13	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
Styrene	<0.50		0.50	ug/L			02/25/22 17:13	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 17:13	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 17:13	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 17:13	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 17:13	1
Toluene	<0.50		0.50	ug/L			02/25/22 17:13	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 17:13	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 17:13	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 17:13	1
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 17:13	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 17:13	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 17:13	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 17:13	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 17:13	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 17:13	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	75		70 - 130		02/25/22 17:13	1
1,2-Dichloroethane-d4 (Surr)	105		70 - 130		02/25/22 17:13	1
4-Bromofluorobenzene (Surr)	82		70 - 130		02/25/22 17:13	1
Toluene-d8 (Surr)	100		70 - 130		02/25/22 17:13	1

Client Sample ID: 202202180212

Lab Sample ID: 810-15928-5

Date Collected: 02/16/22 11:45

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	45.6		0.5000	ug/L			02/25/22 17:36	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 17:36	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 17:36	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 17:36	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 17:36	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 17:36	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 17:36	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 17:36	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 17:36	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 17:36	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 17:36	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180212

Lab Sample ID: 810-15928-5

Date Collected: 02/16/22 11:45

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 17:36	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 17:36	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 17:36	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 17:36	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 17:36	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 17:36	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 17:36	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 17:36	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 17:36	1
Benzene	<0.50		0.50	ug/L			02/25/22 17:36	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
Bromoform	0.51		0.50	ug/L			02/25/22 17:36	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 17:36	1
Bromodichloromethane	12		0.50	ug/L			02/25/22 17:36	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 17:36	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 17:36	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 17:36	1
Chloroform	28		0.50	ug/L			02/25/22 17:36	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 17:36	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 17:36	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 17:36	1
Dibromochloromethane	5.1		0.50	ug/L			02/25/22 17:36	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 17:36	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 17:36	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 17:36	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 17:36	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 17:36	1
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 17:36	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 17:36	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 17:36	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
Styrene	<0.50		0.50	ug/L			02/25/22 17:36	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 17:36	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 17:36	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 17:36	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 17:36	1
Toluene	<0.50		0.50	ug/L			02/25/22 17:36	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 17:36	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 17:36	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 17:36	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180212

Date Collected: 02/16/22 11:45

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-5

Matrix: Drinking Water

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 17:36	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 17:36	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 17:36	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 17:36	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 17:36	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 17:36	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	78		70 - 130		02/25/22 17:36	1
1,2-Dichloroethane-d4 (Surr)	107		70 - 130		02/25/22 17:36	1
4-Bromofluorobenzene (Surr)	82		70 - 130		02/25/22 17:36	1
Toluene-d8 (Surr)	100		70 - 130		02/25/22 17:36	1

Client Sample ID: 202202180213

Date Collected: 02/16/22 10:50

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-6

Matrix: Drinking Water

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	32.80		0.5000	ug/L			02/25/22 17:59	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 17:59	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 17:59	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 17:59	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 17:59	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 17:59	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 17:59	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 17:59	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 17:59	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 17:59	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 17:59	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 17:59	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 17:59	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 17:59	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 17:59	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 17:59	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 17:59	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 17:59	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 17:59	1
Benzene	<0.50		0.50	ug/L			02/25/22 17:59	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 17:59	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180213

Lab Sample ID: 810-15928-6

Date Collected: 02/16/22 10:50

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Bromoform	<0.50		0.50	ug/L			02/25/22 17:59	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 17:59	1
Bromodichloromethane	9.6		0.50	ug/L			02/25/22 17:59	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 17:59	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 17:59	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 17:59	1
Chloroform	19		0.50	ug/L			02/25/22 17:59	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 17:59	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 17:59	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 17:59	1
Dibromochloromethane	4.2		0.50	ug/L			02/25/22 17:59	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 17:59	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 17:59	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 17:59	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 17:59	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 17:59	1
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 17:59	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 17:59	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 17:59	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
Styrene	<0.50		0.50	ug/L			02/25/22 17:59	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 17:59	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 17:59	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 17:59	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 17:59	1
Toluene	<0.50		0.50	ug/L			02/25/22 17:59	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 17:59	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 17:59	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 17:59	1
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 17:59	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 17:59	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 17:59	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 17:59	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 17:59	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 17:59	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	77		70 - 130		02/25/22 17:59	1
1,2-Dichloroethane-d4 (Surr)	102		70 - 130		02/25/22 17:59	1
4-Bromofluorobenzene (Surr)	78		70 - 130		02/25/22 17:59	1
Toluene-d8 (Surr)	99		70 - 130		02/25/22 17:59	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180214

Lab Sample ID: 810-15928-7

Date Collected: 02/16/22 08:50

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	16.80		0.5000	ug/L			02/25/22 18:23	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 18:23	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 18:23	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 18:23	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 18:23	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 18:23	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 18:23	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 18:23	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 18:23	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 18:23	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 18:23	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 18:23	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 18:23	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 18:23	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 18:23	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 18:23	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 18:23	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 18:23	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 18:23	1
Benzene	<0.50		0.50	ug/L			02/25/22 18:23	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
Bromoform	<0.50		0.50	ug/L			02/25/22 18:23	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 18:23	1
Bromodichloromethane	5.9		0.50	ug/L			02/25/22 18:23	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 18:23	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 18:23	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 18:23	1
Chloroform	7.5		0.50	ug/L			02/25/22 18:23	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 18:23	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 18:23	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 18:23	1
Dibromochloromethane	3.4		0.50	ug/L			02/25/22 18:23	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 18:23	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 18:23	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 18:23	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 18:23	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 18:23	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180214

Lab Sample ID: 810-15928-7

Date Collected: 02/16/22 08:50

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 18:23	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 18:23	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 18:23	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
Styrene	<0.50		0.50	ug/L			02/25/22 18:23	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 18:23	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 18:23	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 18:23	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 18:23	1
Toluene	<0.50		0.50	ug/L			02/25/22 18:23	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 18:23	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 18:23	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 18:23	1
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 18:23	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 18:23	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 18:23	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 18:23	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 18:23	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 18:23	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	75		70 - 130		02/25/22 18:23	1
1,2-Dichloroethane-d4 (Surr)	102		70 - 130		02/25/22 18:23	1
4-Bromofluorobenzene (Surr)	80		70 - 130		02/25/22 18:23	1
Toluene-d8 (Surr)	97		70 - 130		02/25/22 18:23	1

Client Sample ID: 202202180215

Lab Sample ID: 810-15928-8

Date Collected: 02/16/22 09:10

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	16.60		0.5000	ug/L			02/25/22 18:46	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 18:46	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 18:46	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 18:46	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 18:46	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 18:46	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 18:46	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 18:46	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 18:46	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 18:46	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 18:46	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180215

Lab Sample ID: 810-15928-8

Date Collected: 02/16/22 09:10

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 18:46	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 18:46	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 18:46	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 18:46	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 18:46	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 18:46	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 18:46	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 18:46	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 18:46	1
Benzene	<0.50		0.50	ug/L			02/25/22 18:46	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
Bromoform	<0.50		0.50	ug/L			02/25/22 18:46	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 18:46	1
Bromodichloromethane	5.8		0.50	ug/L			02/25/22 18:46	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 18:46	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 18:46	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 18:46	1
Chloroform	7.4		0.50	ug/L			02/25/22 18:46	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 18:46	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 18:46	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 18:46	1
Dibromochloromethane	3.4		0.50	ug/L			02/25/22 18:46	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 18:46	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 18:46	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 18:46	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 18:46	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 18:46	1
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 18:46	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 18:46	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 18:46	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
Styrene	<0.50		0.50	ug/L			02/25/22 18:46	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 18:46	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 18:46	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 18:46	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 18:46	1
Toluene	<0.50		0.50	ug/L			02/25/22 18:46	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 18:46	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 18:46	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 18:46	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180215

Date Collected: 02/16/22 09:10

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-8

Matrix: Drinking Water

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 18:46	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 18:46	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 18:46	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 18:46	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 18:46	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 18:46	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	77		70 - 130		02/25/22 18:46	1
1,2-Dichloroethane-d4 (Surr)	103		70 - 130		02/25/22 18:46	1
4-Bromofluorobenzene (Surr)	81		70 - 130		02/25/22 18:46	1
Toluene-d8 (Surr)	98		70 - 130		02/25/22 18:46	1

Client Sample ID: 202202180216

Date Collected: 02/16/22 10:30

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-9

Matrix: Drinking Water

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	20.20		0.5000	ug/L			02/25/22 19:09	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 19:09	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 19:09	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 19:09	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 19:09	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 19:09	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 19:09	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 19:09	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 19:09	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 19:09	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 19:09	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 19:09	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 19:09	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 19:09	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 19:09	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 19:09	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 19:09	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 19:09	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 19:09	1
Benzene	<0.50		0.50	ug/L			02/25/22 19:09	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 19:09	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180216

Lab Sample ID: 810-15928-9

Date Collected: 02/16/22 10:30

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Bromoform	<0.50		0.50	ug/L			02/25/22 19:09	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 19:09	1
Bromodichloromethane	6.9		0.50	ug/L			02/25/22 19:09	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 19:09	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 19:09	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 19:09	1
Chloroform	9.5		0.50	ug/L			02/25/22 19:09	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 19:09	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 19:09	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 19:09	1
Dibromochloromethane	3.8		0.50	ug/L			02/25/22 19:09	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 19:09	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 19:09	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 19:09	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 19:09	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 19:09	1
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 19:09	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 19:09	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 19:09	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
Styrene	<0.50		0.50	ug/L			02/25/22 19:09	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 19:09	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 19:09	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 19:09	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 19:09	1
Toluene	<0.50		0.50	ug/L			02/25/22 19:09	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 19:09	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 19:09	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 19:09	1
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 19:09	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 19:09	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 19:09	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 19:09	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 19:09	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 19:09	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	77		70 - 130		02/25/22 19:09	1
1,2-Dichloroethane-d4 (Surr)	105		70 - 130		02/25/22 19:09	1
4-Bromofluorobenzene (Surr)	83		70 - 130		02/25/22 19:09	1
Toluene-d8 (Surr)	98		70 - 130		02/25/22 19:09	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180217

Lab Sample ID: 810-15928-10

Date Collected: 02/16/22 08:30

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	20.90		0.5000	ug/L			02/25/22 19:32	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 19:32	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 19:32	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 19:32	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 19:32	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 19:32	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 19:32	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 19:32	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 19:32	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 19:32	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 19:32	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 19:32	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 19:32	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 19:32	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 19:32	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 19:32	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 19:32	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 19:32	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 19:32	1
Benzene	<0.50		0.50	ug/L			02/25/22 19:32	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
Bromoform	<0.50		0.50	ug/L			02/25/22 19:32	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 19:32	1
Bromodichloromethane	7.3		0.50	ug/L			02/25/22 19:32	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 19:32	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 19:32	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 19:32	1
Chloroform	9.4		0.50	ug/L			02/25/22 19:32	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 19:32	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 19:32	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 19:32	1
Dibromochloromethane	4.2		0.50	ug/L			02/25/22 19:32	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 19:32	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 19:32	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 19:32	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 19:32	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 19:32	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180217

Date Collected: 02/16/22 08:30

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-10

Matrix: Drinking Water

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 19:32	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 19:32	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 19:32	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
Styrene	<0.50		0.50	ug/L			02/25/22 19:32	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 19:32	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 19:32	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 19:32	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 19:32	1
Toluene	<0.50		0.50	ug/L			02/25/22 19:32	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 19:32	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 19:32	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 19:32	1
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 19:32	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 19:32	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 19:32	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 19:32	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 19:32	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 19:32	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	77		70 - 130		02/25/22 19:32	1
1,2-Dichloroethane-d4 (Surr)	105		70 - 130		02/25/22 19:32	1
4-Bromofluorobenzene (Surr)	84		70 - 130		02/25/22 19:32	1
Toluene-d8 (Surr)	98		70 - 130		02/25/22 19:32	1

Client Sample ID: 202202180218

Date Collected: 02/16/22 12:00

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-11

Matrix: Drinking Water

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	36.50		0.5000	ug/L			02/25/22 19:55	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 19:55	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 19:55	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 19:55	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 19:55	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 19:55	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 19:55	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 19:55	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 19:55	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 19:55	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 19:55	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180218

Lab Sample ID: 810-15928-11

Date Collected: 02/16/22 12:00

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 19:55	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 19:55	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 19:55	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 19:55	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 19:55	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 19:55	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 19:55	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 19:55	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 19:55	1
Benzene	<0.50		0.50	ug/L			02/25/22 19:55	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
Bromoform	<0.50		0.50	ug/L			02/25/22 19:55	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 19:55	1
Bromodichloromethane	9.8		0.50	ug/L			02/25/22 19:55	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 19:55	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 19:55	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 19:55	1
Chloroform	22		0.50	ug/L			02/25/22 19:55	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 19:55	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 19:55	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 19:55	1
Dibromochloromethane	4.7		0.50	ug/L			02/25/22 19:55	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 19:55	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 19:55	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 19:55	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 19:55	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 19:55	1
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 19:55	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 19:55	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 19:55	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
Styrene	<0.50		0.50	ug/L			02/25/22 19:55	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 19:55	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 19:55	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 19:55	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 19:55	1
Toluene	<0.50		0.50	ug/L			02/25/22 19:55	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 19:55	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 19:55	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 19:55	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180218

Date Collected: 02/16/22 12:00

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-11

Matrix: Drinking Water

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 19:55	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 19:55	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 19:55	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 19:55	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 19:55	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 19:55	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	76		70 - 130		02/25/22 19:55	1
1,2-Dichloroethane-d4 (Surr)	106		70 - 130		02/25/22 19:55	1
4-Bromofluorobenzene (Surr)	81		70 - 130		02/25/22 19:55	1
Toluene-d8 (Surr)	98		70 - 130		02/25/22 19:55	1

Client Sample ID: 202202180219

Date Collected: 02/16/22 08:30

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-12

Matrix: Drinking Water

PWSID Number: CA4410014

Method: 524.2 - Total Trihalomethanes

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Trihalomethanes, Total	<0.5000		0.5000	ug/L			02/25/22 20:18	1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 20:18	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 20:18	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 20:18	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 20:18	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 20:18	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 20:18	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 20:18	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 20:18	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 20:18	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 20:18	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 20:18	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 20:18	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 20:18	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 20:18	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 20:18	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 20:18	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 20:18	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 20:18	1
Benzene	<0.50		0.50	ug/L			02/25/22 20:18	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 20:18	1

Eurofins Eaton South Bend

Client Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180219

Lab Sample ID: 810-15928-12

Date Collected: 02/16/22 08:30

Matrix: Drinking Water

Date Received: 02/24/22 10:15

PWSID Number: CA4410014

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Bromoform	<0.50		0.50	ug/L			02/25/22 20:18	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 20:18	1
Bromodichloromethane	<0.50		0.50	ug/L			02/25/22 20:18	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 20:18	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 20:18	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 20:18	1
Chloroform	<0.50		0.50	ug/L			02/25/22 20:18	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 20:18	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 20:18	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 20:18	1
Dibromochloromethane	<0.50		0.50	ug/L			02/25/22 20:18	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 20:18	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 20:18	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 20:18	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 20:18	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 20:18	1
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
Naphthalene	<0.50		0.50	ug/L			02/25/22 20:18	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 20:18	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 20:18	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
Styrene	<0.50		0.50	ug/L			02/25/22 20:18	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 20:18	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 20:18	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 20:18	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 20:18	1
Toluene	<0.50		0.50	ug/L			02/25/22 20:18	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 20:18	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 20:18	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 20:18	1
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 20:18	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 20:18	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 20:18	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 20:18	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 20:18	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 20:18	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	79		70 - 130		02/25/22 20:18	1
1,2-Dichloroethane-d4 (Surr)	108		70 - 130		02/25/22 20:18	1
4-Bromofluorobenzene (Surr)	81		70 - 130		02/25/22 20:18	1
Toluene-d8 (Surr)	97		70 - 130		02/25/22 20:18	1

Eurofins Eaton South Bend

Surrogate Summary

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		DCZ (70-130)	DCA (70-130)	BFB (70-130)	TOL (70-130)
810-15928-1	202202180208	78	103	84	100
810-15928-2	202202180209	77	104	80	100
810-15928-3	202202180210	79	106	81	100
810-15928-4	202202180211	75	105	82	100
810-15928-5	202202180212	78	107	82	100
810-15928-6	202202180213	77	102	78	99
810-15928-7	202202180214	75	102	80	97
810-15928-8	202202180215	77	103	81	98
810-15928-9	202202180216	77	105	83	98
810-15928-10	202202180217	77	105	84	98
810-15928-11	202202180218	76	106	81	98
810-15928-12	202202180219	79	108	81	97
MB 810-13693/5	Method Blank	75	102	81	99

Surrogate Legend

DCZ = 1,2-Dichlorobenzene-d4 (Surr)

DCA = 1,2-Dichloroethane-d4 (Surr)

BFB = 4-Bromofluorobenzene (Surr)

TOL = Toluene-d8 (Surr)

QC Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Method: 524.2 - Volatile Organic Compounds (GC/MS)

Lab Sample ID: MB 810-13693/5

Matrix: Drinking Water

Analysis Batch: 13693

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1,1,1,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 15:26	1
1,1,1-Trichloroethane	<0.50		0.50	ug/L			02/25/22 15:26	1
1,1,2,2-Tetrachloroethane	<0.50		0.50	ug/L			02/25/22 15:26	1
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.50		0.50	ug/L			02/25/22 15:26	1
1,1,2-Trichloroethane	<0.50		0.50	ug/L			02/25/22 15:26	1
1,1-Dichloroethane	<0.50		0.50	ug/L			02/25/22 15:26	1
1,1-Dichloroethene	<0.50		0.50	ug/L			02/25/22 15:26	1
1,1-Dichloropropene	<0.50		0.50	ug/L			02/25/22 15:26	1
1,2,3-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
1,2,3-Trichloropropane	<0.50		0.50	ug/L			02/25/22 15:26	1
1,2,4-Trichlorobenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
1,2,4-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
1,2-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
1,2-Dichloroethane	<0.50		0.50	ug/L			02/25/22 15:26	1
1,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 15:26	1
1,3,5-Trimethylbenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
1,3-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
1,3-Dichloropropane	<0.50		0.50	ug/L			02/25/22 15:26	1
1,4-Dichlorobenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
2,2-Dichloropropane	<0.50		0.50	ug/L			02/25/22 15:26	1
2-Butanone (MEK)	<5.0		5.0	ug/L			02/25/22 15:26	1
2-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 15:26	1
4-Chlorotoluene	<0.50		0.50	ug/L			02/25/22 15:26	1
4-Isopropyltoluene	<0.50		0.50	ug/L			02/25/22 15:26	1
4-Methyl-2-pentanone (MIBK)	<2.0		2.0	ug/L			02/25/22 15:26	1
Benzene	<0.50		0.50	ug/L			02/25/22 15:26	1
Bromobenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
Bromoform	<0.50		0.50	ug/L			02/25/22 15:26	1
Bromomethane	<0.50		0.50	ug/L			02/25/22 15:26	1
Bromodichloromethane	<0.50		0.50	ug/L			02/25/22 15:26	1
Carbon disulfide	<5.0		5.0	ug/L			02/25/22 15:26	1
Carbon tetrachloride	<0.50		0.50	ug/L			02/25/22 15:26	1
Chlorobenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
Chloroethane	<0.50		0.50	ug/L			02/25/22 15:26	1
Chloroform	<0.50		0.50	ug/L			02/25/22 15:26	1
Chloromethane	<0.50		0.50	ug/L			02/25/22 15:26	1
cis-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 15:26	1
cis-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 15:26	1
Dibromochloromethane	<0.50		0.50	ug/L			02/25/22 15:26	1
Dibromomethane	<0.50		0.50	ug/L			02/25/22 15:26	1
Dichlorodifluoromethane	<0.50		0.50	ug/L			02/25/22 15:26	1
Isopropyl ether	<0.50		0.50	ug/L			02/25/22 15:26	1
Ethylbenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
Hexachlorobutadiene	<0.50		0.50	ug/L			02/25/22 15:26	1
Isopropylbenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
Methyl-tert-butyl Ether (MTBE)	<0.50		0.50	ug/L			02/25/22 15:26	1
n-Butylbenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
N-Propylbenzene	<0.50		0.50	ug/L			02/25/22 15:26	1

Eurofins Eaton South Bend

QC Sample Results

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Method: 524.2 - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MB 810-13693/5

Matrix: Drinking Water

Analysis Batch: 13693

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Naphthalene	<0.50		0.50	ug/L			02/25/22 15:26	1
o-Xylene	<0.50		0.50	ug/L			02/25/22 15:26	1
m-Xylene & p-Xylene	<0.50		0.50	ug/L			02/25/22 15:26	1
sec-Butylbenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
Styrene	<0.50		0.50	ug/L			02/25/22 15:26	1
t-Amyl methyl ether	<3.0		3.0	ug/L			02/25/22 15:26	1
Ethyl tert-butyl ether	<3.0		3.0	ug/L			02/25/22 15:26	1
tert-Butylbenzene	<0.50		0.50	ug/L			02/25/22 15:26	1
Tetrachloroethene	<0.50		0.50	ug/L			02/25/22 15:26	1
Toluene	<0.50		0.50	ug/L			02/25/22 15:26	1
trans-1,2-Dichloroethylene	<0.50		0.50	ug/L			02/25/22 15:26	1
trans-1,3-Dichloropropylene	<0.50		0.50	ug/L			02/25/22 15:26	1
Trichloroethylene	<0.50		0.50	ug/L			02/25/22 15:26	1
Trichlorofluoromethane	<0.50		0.50	ug/L			02/25/22 15:26	1
Vinyl chloride	<0.20		0.20	ug/L			02/25/22 15:26	1
Bromoethane	<0.50		0.50	ug/L			02/25/22 15:26	1
Bromochloromethane	<0.50		0.50	ug/L			02/25/22 15:26	1
Dichloromethane	<0.50		0.50	ug/L			02/25/22 15:26	1
Xylenes, Total	<0.50		0.50	ug/L			02/25/22 15:26	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichlorobenzene-d4 (Surr)	75		70 - 130		02/25/22 15:26	1
1,2-Dichloroethane-d4 (Surr)	102		70 - 130		02/25/22 15:26	1
4-Bromofluorobenzene (Surr)	81		70 - 130		02/25/22 15:26	1
Toluene-d8 (Surr)	99		70 - 130		02/25/22 15:26	1

QC Association Summary

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

GC/MS VOA

Analysis Batch: 13693

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-15928-1	202202180208	Total/NA	Drinking Water	524.2	
810-15928-2	202202180209	Total/NA	Drinking Water	524.2	
810-15928-3	202202180210	Total/NA	Drinking Water	524.2	
810-15928-4	202202180211	Total/NA	Drinking Water	524.2	
810-15928-5	202202180212	Total/NA	Drinking Water	524.2	
810-15928-6	202202180213	Total/NA	Drinking Water	524.2	
810-15928-7	202202180214	Total/NA	Drinking Water	524.2	
810-15928-8	202202180215	Total/NA	Drinking Water	524.2	
810-15928-9	202202180216	Total/NA	Drinking Water	524.2	
810-15928-10	202202180217	Total/NA	Drinking Water	524.2	
810-15928-11	202202180218	Total/NA	Drinking Water	524.2	
810-15928-12	202202180219	Total/NA	Drinking Water	524.2	
MB 810-13693/5	Method Blank	Total/NA	Drinking Water	524.2	

Analysis Batch: 13978

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-15928-1	202202180208	Total/NA	Drinking Water	524.2	
810-15928-2	202202180209	Total/NA	Drinking Water	524.2	
810-15928-3	202202180210	Total/NA	Drinking Water	524.2	
810-15928-4	202202180211	Total/NA	Drinking Water	524.2	
810-15928-5	202202180212	Total/NA	Drinking Water	524.2	
810-15928-6	202202180213	Total/NA	Drinking Water	524.2	
810-15928-7	202202180214	Total/NA	Drinking Water	524.2	
810-15928-8	202202180215	Total/NA	Drinking Water	524.2	
810-15928-9	202202180216	Total/NA	Drinking Water	524.2	
810-15928-10	202202180217	Total/NA	Drinking Water	524.2	
810-15928-11	202202180218	Total/NA	Drinking Water	524.2	
810-15928-12	202202180219	Total/NA	Drinking Water	524.2	

Lab Chronicle

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180208

Date Collected: 02/16/22 11:20

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-1

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 16:03	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 16:03	CM	EA SB

Client Sample ID: 202202180209

Date Collected: 02/16/22 09:30

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-2

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 16:26	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 16:26	CM	EA SB

Client Sample ID: 202202180210

Date Collected: 02/16/22 10:10

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-3

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 16:50	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 16:50	CM	EA SB

Client Sample ID: 202202180211

Date Collected: 02/16/22 09:50

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-4

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 17:13	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 17:13	CM	EA SB

Client Sample ID: 202202180212

Date Collected: 02/16/22 11:45

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-5

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 17:36	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 17:36	CM	EA SB

Client Sample ID: 202202180213

Date Collected: 02/16/22 10:50

Date Received: 02/24/22 10:15

Lab Sample ID: 810-15928-6

Matrix: Drinking Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 17:59	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 17:59	CM	EA SB

Eurofins Eaton South Bend

Lab Chronicle

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Client Sample ID: 202202180214

Lab Sample ID: 810-15928-7

Date Collected: 02/16/22 08:50

Matrix: Drinking Water

Date Received: 02/24/22 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 18:23	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 18:23	CM	EA SB

Client Sample ID: 202202180215

Lab Sample ID: 810-15928-8

Date Collected: 02/16/22 09:10

Matrix: Drinking Water

Date Received: 02/24/22 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 18:46	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 18:46	CM	EA SB

Client Sample ID: 202202180216

Lab Sample ID: 810-15928-9

Date Collected: 02/16/22 10:30

Matrix: Drinking Water

Date Received: 02/24/22 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 19:09	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 19:09	CM	EA SB

Client Sample ID: 202202180217

Lab Sample ID: 810-15928-10

Date Collected: 02/16/22 08:30

Matrix: Drinking Water

Date Received: 02/24/22 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 19:32	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 19:32	CM	EA SB

Client Sample ID: 202202180218

Lab Sample ID: 810-15928-11

Date Collected: 02/16/22 12:00

Matrix: Drinking Water

Date Received: 02/24/22 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 19:55	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 19:55	CM	EA SB

Client Sample ID: 202202180219

Lab Sample ID: 810-15928-12

Date Collected: 02/16/22 08:30

Matrix: Drinking Water

Date Received: 02/24/22 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	524.2		1	13693	02/25/22 20:18	CM	EA SB
Total/NA	Analysis	524.2		1	13978	02/25/22 20:18	CM	EA SB

Laboratory References:

EA SB = Eurofins Eaton South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Eurofins Eaton South Bend

Accreditation/Certification Summary

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Laboratory: Eurofins Eaton South Bend

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
California	State	2920	06-30-22

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
524.2		Drinking Water	1,1-Dichloropropene
524.2		Drinking Water	1,2,3-Trichloropropane
524.2		Drinking Water	1,3-Dichloropropane
524.2		Drinking Water	2,2-Dichloropropane
524.2		Drinking Water	2-Butanone (MEK)
524.2		Drinking Water	4-Isopropyltoluene
524.2		Drinking Water	Bromobenzene
524.2		Drinking Water	Bromochloromethane
524.2		Drinking Water	Bromoethane
524.2		Drinking Water	Bromomethane
524.2		Drinking Water	Chloroethane
524.2		Drinking Water	Chloromethane
524.2		Drinking Water	Dibromomethane
524.2		Drinking Water	Hexachlorobutadiene
524.2		Drinking Water	Isopropyl ether
524.2		Drinking Water	Trihalomethanes, Total

Method Summary

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Method	Method Description	Protocol	Laboratory
524.2	Total Trihalomethanes	EPA-DW	EA SB
524.2	Volatile Organic Compounds (GC/MS)	EPA-DW	EA SB

Protocol References:

EPA-DW = "Methods For The Determination Of Organic Compounds In Drinking Water", EPA/600/4-88/039, December 1988 And Its Supplements.

Laboratory References:

EA SB = Eurofins Eaton South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Sample Summary

Client: Eurofins Eaton Analytical
Project/Site: 988302

Job ID: 810-15928-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
810-15928-1	202202180208	Drinking Water	02/16/22 11:20	02/24/22 10:15	CA4410014
810-15928-2	202202180209	Drinking Water	02/16/22 09:30	02/24/22 10:15	CA4410014
810-15928-3	202202180210	Drinking Water	02/16/22 10:10	02/24/22 10:15	CA4410014
810-15928-4	202202180211	Drinking Water	02/16/22 09:50	02/24/22 10:15	CA4410014
810-15928-5	202202180212	Drinking Water	02/16/22 11:45	02/24/22 10:15	CA4410014
810-15928-6	202202180213	Drinking Water	02/16/22 10:50	02/24/22 10:15	CA4410014
810-15928-7	202202180214	Drinking Water	02/16/22 08:50	02/24/22 10:15	CA4410014
810-15928-8	202202180215	Drinking Water	02/16/22 09:10	02/24/22 10:15	CA4410014
810-15928-9	202202180216	Drinking Water	02/16/22 10:30	02/24/22 10:15	CA4410014
810-15928-10	202202180217	Drinking Water	02/16/22 08:30	02/24/22 10:15	CA4410014
810-15928-11	202202180218	Drinking Water	02/16/22 12:00	02/24/22 10:15	CA4410014
810-15928-12	202202180219	Drinking Water	02/16/22 08:30	02/24/22 10:15	CA4410014

Submittal Form

*REPORTING REQUIREMENTS: Do Not Combine Reports with any other samples submitted under different Folder Numbers!
Report & Invoice must have the Folder # 988302 Job # 1000014

Report all quality control data according to Method. Include dates analyzed. Date extracted (if extracted) and Method reference on the report.
Results must have Complete data & QC with Approval Signature.

Ship To:
Eurofins Eaton Analytical
110 South Hill Street
South Bend, IN 46617-2702

Phone: 800-332-4345 Fax: 574-233-8207

Folder #:
988302

Report Due:
03/03/2022

Reports: Jackie Contreras Sub-Contracting Administrator
EMAIL TO: Eaton-MonroviaSubContract@eurofinset.com
Eurofins Eaton Analytical, LLC 750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016
Phone (626) 386-1165 Fax (626) 386-1122
Invoices to: Eurofins Eaton Analytical, LLC
Accounts Payable 2425 New Holland Pike, Lancaster, PA 17605

Provide in each Report the
Specified State Certification # and
Exp Date for requested tests + matrix.
Samples from: CALIFORNIA

Monica is the ASM - 524.2 Monrovia list with Bromoethane
run trip blank



Sample ID 202202180208	Client Sample ID for reference onl 1000 Stewart Dr	Sample Date & Time 02/16/22 1120 DW	Clip Code CA4410014_DST_909	PWSID JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:

Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)
---------------------	--------------------------	---

Sample ID 202202180209	Client Sample ID for reference onl 12788 Hwy 9	Sample Date & Time 02/16/22 0930 DW	Clip Code CA4410014_DST_902	PWSID JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:

Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)
---------------------	--------------------------	---

Sample ID 202202180210	Client Sample ID for reference onl 15030 Hwy 9	Sample Date & Time 02/16/22 1010 DW	Clip Code CA4410014_DST_905	PWSID JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:

Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)
---------------------	--------------------------	---

Client Provided Sample Container

Amber

20°C

Relinquished by: [Signature] Sample Control Date 02/23/22 Time 14:08

Received by: [Signature] Date 2/24/22 Time 10:15

Relinquished by: [Signature] Sample Control Date 2/24/22 Time 10:15

Received by: [Signature] Date 2/24/22 Time 10:15

NOTIFICATION REQUIRED IF RECEIVED OUTSIDE OF 0-6 CELSIUS

An Acknowledgement of Receipt is requested to attn: Jackie Contreras

Sample ID 202202180211	Client Sample ID for reference on! 275 Brook Ln	Sample Date & Time 02/16/22 0950 DW	Clip Code CA4410014_DST_904	PWSID JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:
Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)		
Sample ID 202202180212	Client Sample ID for reference on! 801 Dry Well Rd	Sample Date & Time 02/16/22 1145 DW	Clip Code CA4410014_DST_908	PWSID JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:
Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)		
Sample ID 202202180213	Client Sample ID for reference on! 814 Creek Drive	Sample Date & Time 02/16/22 1050 DW	Clip Code CA4410014_DST_907	PWSID JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:
Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)		
Sample ID 202202180214	Client Sample ID for reference on! 815 Forest Drive	Sample Date & Time 02/16/22 0850 DW	Clip Code CA4410014_DST_901	PWSID JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:
Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)		
Sample ID 202202180215	Client Sample ID for reference on! Clear Creek and Alameda	Sample Date & Time 02/16/22 0910 DW	Clip Code CA4410014_DST_903	PWSID JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:
Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)		
Sample ID 202202180216	Client Sample ID for reference on! River and Crescent Dr	Sample Date & Time 02/16/22 1030 DW	Clip Code CA4410014_DST_906	PWSID JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:
Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)		

Client Provided Sample Container

2.0°C

NOTIFICATION REQUIRED IF RECEIVED OUTSIDE OF 0-6 CELSIUS

An Acknowledgement of Receipt is requested to attn: Jackie Contreras

Relinquished by: *[Signature]* Date 02/23/22 Time 14:08

Received by: *[Signature]* Date 2/24/22 Time 10:15

Relinquished by: Sample Control Date Time

Received by: *[Signature]* Date Time

Page 8 of 8 5522422

750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016 Tel (626) 386-1100 Fax (866) 988-3757 www.EurofinsUS.com/Eaton

Page 37 of 39

3/7/2022

Sample ID 202202180217	Client Sample ID for reference onl Riverside and Middle		Sample Date & Time 02/16/22 0830 DW	Clip Code CA4410014_DST_911	PWSID	JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:		
Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)				
Sample ID 202202180218	Client Sample ID for reference onl Sylvan and Redwood		Sample Date & Time 02/16/22 1200 DW	Clip Code CA4410014_DST_910	PWSID	JLS
Sample type: RT	Sample Event:	Facility ID:	Sample Point ID:	Static ID:		
Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)				
Sample ID 202202180219	Client Sample ID for reference onl Trip Blank C		Sample Date & Time 02/16/22 0830 DW	Clip Code	PWSID	JLS
Sample type:	Sample Event:	Facility ID:	Sample Point ID:	Static ID:		
Method EPA 524.2	Prep Method EPA 524.2	Analysis Requested VOC EPA 524.2 Monrovia List (2)				

Client Provided Sample Container

2.0°C

Relinquished by: [Signature] Sample Control B. RETNER Date 02-23-22 Time 14:08

NOTIFICATION REQUIRED IF RECEIVED OUTSIDE OF 0-6 CELSIUS

Received by: [Signature] Date 2-24-22 Time 1015

An Acknowledgement of Receipt is requested to attn: Jackie Contreras

Relinquished by: [Signature] Date 2-24-22 Time 1015

Sample Control

Received by: [Signature] Date 2-24-22 Time 1015

Page 38 of 39

750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016 Tel (626) 386-1100 Fax (866) 988-3757 www.EurofinsUS.com/Eaton

Page 38 of 39

3/7/2022

Login Sample Receipt Checklist

Client: Eurofins Eaton Analytical

Job Number: 810-15928-1

Login Number: 15928

List Number: 1

Creator: Spurgeon, Sheri

List Source: Eurofins Eaton South Bend

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	False	Client provided containers